Emergency Bag / AED / O2 Checklist

(Retain Checklist for 3 years.)

*To be checked daily. (Please make notation for weekends/holidays)

Month/Year:	Unit/Dept:

Date and Initial	AED Check for "OK"	AED Electrodes 2 packages in pocket Check Expiration Date	Red Plastic tamper-resistant lock on Emergency Bag Intact. Lock Number	Bag with Pocket mask, gloves & razor (on AED) Plastic tamper- resistant lock intact	Oxygen Tank Valve in OFF position Gauge reads FULL	Forms inside pocket CPR Flowsheet Evaluation / Debriefing Form
1						
2						
3						
4						
5						
6						
7						
8						
9						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						
31						